

# Implement of Medical Application on Wireless Board Band Network (Wi-MAX) System

K.NOIMANEE\*, S.NOIMANEE\*\*, S. TANTISATIRAPONG\*\*\*

\*Department of Computer Engineering, Faculty of Engineering, Rajamangala University of Technology Lanna, 123 Hauy Keaw Rd., Chiang Mai 50200, THAILAND.

\*\*Department of Forensic Science, Faculty of Liberal Arts and Science, Kasetsart University, Kamphaeng Campus, 1 Moo 6 Kamphaeng Saen District Nakhon Pathom 73140, THAILAND.

\*\*\*Department of Biomedical Engineering, Faculty of Engineering, Srinakharinwirot University, 63 Moo 7, Ongkharak, Nakhon Nayok 26120, THAILAND.

E-mail: all\_of\_none@hotmail.com, dr.suranan@gmail.com, beau\_suchada@yahoo.com

**Abstract:** - This paper present the functional parts of a wireless board band system (Wi-MAX), its basic operation and medical application in Thailand. Our research explain how it can use base stations to provide high speed data connections that can be used for voice, medical data and medical consult video services (MCVS) to distances of over 20 km. Anywhere, the project aimed to research develop of wireless broadband network for medical applications of Medical Center, Prince Maha Chakri Sirindhorn's, HRH Princess Siridhorn, Amphor Ongklaruck, Nakhon Nayok Province, Thailand. This project is integrating other sciences, biomedical engineering, telecommunications engineering, electronic engineering and computing technology. Includes both direct medical and so on. After this project after the completion of phase 1, then we have to focus on medical care, community living around the area radius of the radio wave propagation in the district that spreads to the 2.500-2.520 GHz frequency range in Thailand by the standard of broadband wireless technology. This project involved 2 Subprojects, namely: Firstly, is development of multi patient monitoring on wireless board band technology around MSMC. And finally subproject is medical information application for medical image data in MSMC.

**Key-Words:** medical applications, Prince Maha Chakri Sirindhorn's, MCVS, HRH Princess Siridhorn, 2.500-2.520GHz frequency range, multi patient monitoring, Wi-MAX, MSMC.

## 1. Introduction

Thailand is today considered to be the top Medical services with a very successful hospital center best service more than 80% of world services share. In the quest to identify the next generation medical service, the hospital center in Nakorn Nayok province chose WiMAX-Thai Biomedical Engineering Research Association (ThaiBME) forum-defined implementation of medical services through a WiMAX technology-to open up another frontier in the medical services. In 2012, the committee ThaiBME group of Executive ThaiBME's staff developed a WiMAX system with National Telecommunication Commission, which states WiMAX technology will be a focus for the future of Thailand Medical Center Services (MSMC), and will be the preferred technology to deliver e-Health, medical consult in the "Mobile WiMAX program". The Mobile-WiMAX program and medical application acceleration projects, which aim to develop medical services and base stations, were initiated under this system. Towards this end, medical center (Prince Maha Chakri Sirindhorn's)

offers only world class services capability, and also an entire service, medical consult trough an WiMAX television, by forming its own system. Forming a complete WiMAX system structure, including 2.5GHz base station, CPE, network elements, physician teams, system integration, medical application, ICT operation and services, is part of the overall strategy adopted by MSMC for development of technology. Cooperating with MSMC's physicians, and leveraging the strengths of its medical services group, Thailand is able to speed up system development, medical services and technology of WiMAX system. WiMAX spectrum with good physical characteristics, which is crucial to network deployment, has been planned and will be released by September 2012 to facilitate the MSMC program and subsequent medical work of WiMAX systems. To promote global interoperability MSMC actively participates in international standards of medical services and promotes international collaborations. Thailand has been working closely with the NTC forum in promoting and sharing physician-Thailand

experiences. Section 2 of this case study describes structure of WiMAX system, Section 3 discusses WiMAX testing on site, Section 4 describes vital-sign measurement through WiMAX network, Section 5 describes antenna design for WiMAX system, Section 6 describes the design of experiment, Section 7 describes WiMAX equipments and software configuration and Section 8 provides conclusions. The material for this paper was provided by Assoc.Prof.Dr. Suranan Noimanee of computer engineering; who is the liaison for NTC of Thailand forum. We coordinated contributions from other members of ThaiBME and Srinakarinwirot University, was prepared by Plannet Communications company Ltd. on behalf of the liaison for WiMAX of NTC forum. Contributors were: Professor Prasit Prapinmongkolkarn Chairman of the National Telecommunication (NTC) of Thailand and Mr. Prapat Ratleardkarn. Anyway, HRH Princess Maha Chakri Sirindhorn Medical Center has revised its plan to proceed. Because of the need to strengthen information systems and basic medical services are available within the Medical Center. Before you take a high-speed wireless broadband WiMAX is set to be released to the medical center's medical services. They tried working on the biomedical engineering researchers project's to bring their medical application. Such as in medical applications, the vital-signs measurement from faculty of engineering, Chiang Mai University, faculty of engineering KMITL and faculty of Science, KMUTL etc,. The medical center has received donation from the National Telecommunications Commission in amount of 1,000,000 USD. Its original purpose was:- Firstly, to give a royal charity dedicated to occasion age of a King Bhumibol Adulyade 80<sup>th</sup> birthday. Secondly, to provide a center for developing and linking information of the patient's data. And services both inside and outside the medical center for development of public health data system. Thirdly, to develop information technology computer network in order to manage the data and knowledge of public health center, clinical network, hospital around the province and other servers. Next, to create a network of research and knowledge management, medical and public health. Finally, to develop information systems in facilitating services for patients at the referral center at the medical center and returned to community raised area. Therefore, to test the feasibility of adopting WiMAX technology to try to establish a new communication channel called telemedicine application for medical research talk face to face method. Called long-distance consultation with in

various medical information. Such as vital-signed and medical images in digital format is the central of the medical center to the health substation, include mobile client unit. And to study their effects such as interference of high speed broadband spectrum with the old system in medical center located. By studying the possibility of bringing the high speed broadband technology (WiMAX). Used to medical services. The medical center as a lead agency in managing the project and is the medical network hub. There was conducted in the following area is completed. They connected via radio frequency to client 2 health substations in Nakorn Nayok province. In this paper, we proposed medical application by WiMAX technology used to create new communication channels to the National Telecommunications Commission. At the end of the period of project. Substance of the report covers issues such as requirements and technical feasibility. Telecommunication network features and components that contain the details of the spectrum, the routes to and use of network calls. Transportation costs in serving as a model to develop in the future. For monitoring and evaluation centers. Physician or medical doctors had been monitoring and evaluation of every project and report results to the Office of National Telecommunications Broadcasting and Telecommunication Commission (NBTC) to, as mentioned above. Detailed results of the project are as follows:- Activities planed: Education and development of medical applications, including vital signs monitoring of patients on the ambulance station. Health Center houses the public that medical doctors are required to determine the radius of the house is surrounded by high-speed wireless broadband (WiMax) of the hospital information system development the medical information system of a medical imaging center. The Medical/analysis under the three sub-projects including. Firstly, Sub-project 1<sup>st</sup>: Development of Multi Patient Monitoring on WiMAX Technology Around MSMC and secondary is Sub-project 2<sup>nd</sup>: Primary Health Care Mobile Doctor Around MSMC.

## 2. System Implementation

### 2.1 Structure of system profile.

A WiMAX technology as certification follow the IEEE 802.16 space interface specifications from IEEE organization. The network specifications of WiMAX hardware product. However, they are being developed internally by the WiMAX of NTC Thailand forum, which include the end-to-end networking specifications and network

interoperability specifications. The Network Working Group within the WiMAX forum is responsible for these network specifications, some of which involve Access Service Network (ASN) control and data plane protocols, ASN profiles, Connectivity Services Network (CSN) mobility support, Authentication, Authorization and Accounting (AAA) interworking with other technologies, and various services such as Location-Based Service (LBS), Multicast and Broadcast Service (MCBCS) etc. In this chapter, I will focus on the overview of WiMAX technology from the air interface perspective.




Figure 1. Fixed and mobile WiMAX first release certification.

Figure 1 presents the aforementioned composition of the current WiMAX technology, commonly referred to as first release profile. Its air interface specifications consist of four related IEEE 802.16 Broadband Wireless Access Standards, that is, IEEE standard 802.16-2004, IEEE standard 802.16-2004/Cor.1-2005, IEEE standard 802.16e-2005 and the IEEE draft standard P802.16-2004/Cor.2. Not all of the optional features defined in these IEEE standards are implemented in WiMAX hardware and tested for certifications. Through extensive technical investigation analysis to build up the best competitive hardware, the WiMAX forum technical working in Thailand group published the first version of WiMAX system first profile in early 2007. The latest published version to date incorporated error fixes and minor corrections without touching the main features selected in the first revision.




Figure 2 Structure of Fixed and Mobile WiMAX system profile.

Figure 2 shown a more detailed view of the construction of the WiMAX system profile is presented from the space interface perspective. The system profile is composed of five sub-profiles, namely, PHY, MAC, radio, duplexing mode and power classes. Even though there are many different combinations of center frequencies and channel bandwidths accommodating different regional spectrum regulations, all Release 1 mobile WiMAX products share the same PHY and MAC features (profiles) and the same duplexing mode which is Time Division Duplex (TDD). In the following, some detailed descriptions of key PHY and MAC features in the mobile WiMAX system profile are offered.

## 2.2 Wi-MAX system description.

Our research within the framework of the joint (NTC) by the study and development of medical applications on the network, high-speed wireless broadband (WiMAX)/how to implement research applications. Researchers from the biomedical engineering research activities at the proposal stage 1, and the Memorandum of Agreement (MOU) between the NTC and the MSMC. The structure in Figure 3 and Figure 4, in various forms and with developments in other parts. Is planned within the framework of first and third sub-projects, working with researchers at the Research Society of Biomedical Engineering to the appropriateness and efficiency of research and development.




Figure 3. Medical applications by connecting to the network to the second health district.




Figure 4. Structure of telemedicine via Wi-MAX

The operation of vital-signs in Figure 4 shows the following:-

1. How to measure the vital signs of patients. Patients must be attached to electrodes on the patient every time a measurement.
2. Signs are based on measurements obtained from the module receiver - radio frequency 2.4 MHz (standard IEEE 802.16.4), which has low power as possible. (Can be adjusted from 5-100 mW, which is to prevent harm to patients and medical equipment).
3. A patient's vital signs will be transmitted to the telemedicine table has been around less than 10 meters.
4. The CPE on the telemedicine table, it sends a signal to the WiMAX CPE to connect to the Medical Center, King of the Chakri Sirindhorn. HRH Princess Maha Chakri Sirindhorn. In this study, in a radius of 8-20 k.m.

This project has created a network of research and knowledge management, medical and public health systems will be used to patient monitoring such as the heart sounds, lung sounds, electrocardiogram via WiMAX network.




Figure 5. Overview of medical application via wireless broadband technology.

Figure 5 is a model to measure vital signs, the medical doctor remotely via teleconference at the tele-consultation and treatment of a doctor, whether in the medical center's size or small health district. Physicians can provide patients via high-speed wireless broadband (WiMAX) has the following functions: -

1. Measuring vital signs of patients. Patients must be attached to electrodes on the patient every time a measurement.
2. Vital signs were measured to the modules to transceiver frequency of 2.4 GHz with low power as possible. (Adjustable from 5-50 mW, which is to prevent harm to patients and medical equipment).
3. A patient's vital-signs will be sent to a nearby table during telemedicine in less than 10 meters.
4. The telemedicine equipment CPE on the table, it sends a signal to the WiMAX CPE to connect to a medical center in this study were within the radius of 80-20 kilometers.
5. The research of this project has created a network of research and knowledge management, medical and public health will bring a variety of patient monitoring systems. Such as heart and lung sound monitoring, electrocardiogram monitoring or other vital-signs monitoring through WiMAX (Prototype shown in Figure 14 and how to measure the ECG through WiMAX system).

### 2.3 Design of Implementation

ECG signal processing system design for long-distance wireless data transmission module for low power, this requires knowledge of many aspects. Namely, education, medical theory for the origin of the ECG using the basic design of electronic circuits, the development of micro-controllers,

communication systems, and development of software in the computer to be able to perform. The research done in the design of the overall system will be structured as shown in Figure 6.




Figure 6. The structure of the signal processing system in the ECG monitoring

Figure 6 is shown the overall structure of the system consists of a portable device for the patient is responsible for monitoring the ECG and transmit signals via a wireless network in the 100 meters to the computer display and storage. Under the analysis of signals for signs of unusual and could signal a warning to doctors who treat over the phone, this system is able to accommodate the connection of the patients were 8 people who use the system. They can be connected to a network computer with a user is an administrator, medical treatment and general personnel.

### 2.3.1 Design of ECG transmitter device through Wi-MAX Backbone.

The design of the ECG of a channel and through a wireless network with a modular transceiver ZigBee is shown in Figure 7 contains the measured signal using the detection signals from the skin of patients with Electrodes, ECG. is represented by the voltage coming out with Differential Amplifier circuit and the voltage of electricity that can be used to study the function of the heart, the signal that it is bandwidth in the range between 1 - 100 Hz. for interference or radio frequency interference, etc. that may come from the patient himself, was then designed to amplify the signal and noise reduction by using the signal of Active device, such as the Operational Amplifier to reduce noise, the signal that it is. The analog signal will be sent to. Number PIC16F675 microcontroller, which is responsible for converting the signals into digital signals with a set of modules, ZigBee IEEE 802.15.4 transceiver to transmit a signal to the computer data center.




Figure 7. The design of ECG monitoring via Wi-MAX block diagram.

### 2.3.2 Computer programs for processing and notification of an emergency.

The hospital's computer server will act as a hub of information processing. Store and check for signs of the patient. It can sending a signal warning to physicians and caregivers in an emergency. It consists of a set of program data for receive the software and data storage and processing of signals.

### 2.3.3 Receiving and storage of ECG data.

Hospital computer server will support the transmission of many patients at the same time. It is a server -based applications developed with Visual Basic working with multi-track processing at the same time in 2 parts. The acting in data management and signal processing in the receiving and storage of this data. It can be displayed in the connection status of each patient. And ECG signal graph of the many patients at the same time. In the other part is connected to an online database, which uses a MySQL database to support the user's connection to the web service to the application data flow diagram is shown in Figure 8.




Figure 8. The Data Flow Diagram of ECG signal processing application

To work “Online” physician will be shows the vital-signs of the patient in real time before you have saved data in the database. And the “Offline” he can call the previous patient's database. In the database design will be the structure shown in Figure 9.




Figure 9. Database of ECG signal processing through WiMAX



Figure 10. Measurement results are displayed in real time via WiMAX.

Figure 10 is experimental device developed ECG monitoring via Wi-MAX which the main program window will display a list of all patients in the database. And linker status in real time and display time data from the received data. Time to end the transmission of data, number of connections that can detect heart rate. Users can select a patient's ECG display in real time.

### 2.3.4 Web Services for connecting with users

PC as a hub or network switch of data, including Apache Web Server Version 2.2.3 PHP Script Language Version 5.1.6 and MySQL Database version 5.0.24a. Web services are web services connected with the development of the programming language PHP. Consists of showing the connection of the patient. The output signal E C G is a graph, most of the details of the patient And in parts of the graph to aid in the diagnosis, analysis, showing details of a data flow diagram in Figure 11.




Figure 11. Data Flow Diagram of web services via WiMAX system and the internet

On the part of the show via computer network, users will be able to show our patients with the same application on server and display real time status of the test results in Figure 12. It shows a web of the ECG signal via Wi-MAX and internet.




Figure 12. Shows the web of the ECG database via Wi-MAX and the Internet



Figure 13. Shows the web of the ECG database via Wi-MAX and the Internet

### 2.3.5 WiMAX Testing on-site at MSMC and Chiang Mai University

The application was developed by a medical team of researchers from Thailand Biomedical Engineering Research Association (ThaiBME), and Assoc.Prof.Dr. Suranan Noimanee Department of Computer Engineering faculty of Engineering Chiang Mai University, Thailand is shows in Figure 14.



Figure 14. ECG&HS equipment prototype with monitoring via Wi-MAX network.

Figure 16 is the shape of the spectrum range from 2.5 GHz antenna (Figure 18), which has studied the interference of the spectrum of the broadband wireless transmission and propagation from the antenna, then this is it. the frequency stability is equal to 2.36 ppm at 27°C and is reduced to 0.003 ppm when the GPS is already unlocked. Which is acceptable. (Default is 2.5 ppm at 30°C and is reduced to 0.003 ppm when the lock on the GPS). The system can take on this project came from the Health and Medical Center. Medical doctor in the area surrounding the medical center, with the cover, it can retrieve data or vital signs of patients and can make a tele-consultation the telemedicine table or Notebook when he removed the patient according to smart phone of physician etc.



Figure 15. Tele-Consultation and physician see patients from around the home.




Figure 16. Radiation pattern of 3-sectors antenna system, the distance between the MSMC to two health sub-district



Figure 17. The 3-sectors Wi-MAX's antenna at Health center and MSMC building.

The MSMC has been supplying, installing, and testing of network devices and connections to Buangsarn and Chumphon health district, where the system has the following main components:-

1. Base Transceiver Station (BTS) devices.
2. Provisioning server devices.
3. EMS server device, such as VoIP server, Video server etc.
4. CPE (Customer Premise Equipment) devices.

**The main function of each device can be explained as follow:-**

1. BTS device serves as a wireless bridge can be send data to the CPE device across the state of the LOS (Line-of-Sight) and NLOS (None-Line-of-Sight) with radius to be covered as far as 8 kilometers, or may be increased up to 32 kilometers if you change the distance of the device to get BTS.

2. Provisioning Server device acts as a device to monitor the implementation of the CPE device to get access to the network. The CPE devices to get access to detailed information of its own Packet to/from the device, so the BTS device will be forwarded to the Provisioning Server make sure that access to work or not. This test can pick up - delivery needs.
3. EMS Server device will check the status of each device is not functioning normally or not. Using the SNMP protocol and status message when the device is working deviated from the norm. This will make management easier.
4. CPE device is a terminal device for use at the local health district, both by combining the functionality of the Transceiver, Modem, and the antenna together (Figure 18) by activating the device CPE is trying to connect BTS equipment for use on networks.



Figure 18. The 3-sectors BTS antenna installation.

There are many type of CPE devices, such as indoor a desktop application (RSU: Residential Subscriber Unit), are active outdoors (OSU: Outdoor Subscriber unit), are installed on the vehicle. (If installed at an ambulance. The physician's can visit patients at home) for use in mobile (MSU: Mobile Subscriber Unit) and the type installed on a portable computer (PC Card). Thus, in analyzing the radio signals (Terrain Analysis) of the BTS (Base Transceiver Station) devices in Figure 16 is derived from the design of the research team, which can be described as follows:-

From the simulation on the Terrain database using Google Earth's software by providing equipment, BTS stuck at a height of about 60 meters

(deck floor 16<sup>th</sup> of the MSMC building) and use the 90 degrees antenna sector and OSU. CPE is the receiver altitude of 2 meters from that which we can detect the signal from each point as shown in Figure 11. The results of the test antenna gain is equal to 13dBd give Rx Level = 61.1 dBm at a distance of 8.67 km is the Path Loss = 120.67 dB Interference by the location of the radio was BuengSan health center and other high voltage equipment. Similarly, the Chumphon health center will be a different test at the 14.80 km distance as shown in Figure 19.




Figure 19. WiMAX radiation pattern testing at Chumphon health center

### 2.3.6 Medical Information Application for Medical Image Data in MSMC

This project is a medical application, medical information and other materials. Such as patient data, medical image data to be transmitted from the medical center to the health district both in height speed. Which is development by Thai Biomedical Research Association. Digital images from x-ray will be stored and processed information to a format that is standardized and reliable. To find and run it later. This is a standard of PACS (Picture Archiving and Communication System). It is a system used to store the medical images and used in the transmission of image data in digital form by using the management data through high-speed wireless broadband computer network. By sending the image format called DICOM. The standards set by the National Electrical Manufacturers Association (NEMA). Intended for publication of medical images such as CT scans, MRI and other medical images. To the same standard will be exchanged during medical or software to be read in a medical patient data/information that comes from a different species or different companies. PACS system is a medical information technology developed initially for use in the department of radiation directly to the surgery because the photos of radiation needed to help diagnose diseases and treat patients. It will allow the doctor to get the latest and fast analysis of radiologists. In addition, my research team will be more applications to process image data into the image data from X-ray seen as a 2-dimensional

conventional display is a 3-dimensional as shown in Figure 80. Which contains information that can help in diagnosis the disease has a more open-ended or even in the future, it may be added as a 4-dimensional (3-D image of the movement).




Figure 20. The 3D images to aid in the diagnosis through WiMAX .

The research team presented the PACS system and image processing system that can prevent data loss and security in the use of a computer network to work today, and WiMAX networks as shown in Figure 81. This system has the ability to use the PACS via web application is also available. PACS system as well as a personal computer running a network of central hospitals or personal computers at health stations. Or you can use the PACS system via the Pocket PC makes it possible to monitor the patient while at home. The system also can keep the equipment safe from hacker with Firewall security. It can define security rules for the system with the MAC and IP Address to be read from the database, and the it is running the RAID system which will prevent data loss in case of hard disk storage device malfunction. For security to the access WiMAX network, my researcher team has designed a device that is connected to the PACS system of medical center must have the characteristics of the packet only one has a code to transmit data each is unique. It changes with time WiMAX devices of other agencies. This may have future access to the network of hospitals medical centers had. In addition, the WiMAX devices to access the User name and Password must be set to match the radius. The system stores the image data for physicians to collect data on the storage system on the network storage devices, the medical imaging equipment such as X-Ray equipment, MRI equipment, they can

store information into categories and to prevent loss due to perform a backup periods of inactivity the system or are used. Research team proposed system to the snoring sounds of a patient via system then the it can to check stored in data files and it can be opened from the PACS system has made medical treatment can be analyzed snoring sounds of the patient.

### 2.3.7 PACS system in MSMC Medical center.

The PACS system must to connected with other system in the medical center, such as Hospital Information System (HIS) and Scanner such as X-ray digital, Ultra sound, CT Scanner which interacts with the LAN as shown in Figure 21.




Figure 21. Structure of PACS in MSMC

Figure 21 is shown physicians are observation medical images sent from other hospitals to obtain consulting through Wi-MAX system with the PACS. The X-Ray digital, Ultra sound, there will be a small number of images each file has a size of about 500-1500 Kbyte. But the CT scanner will have a series of 100-200 files and file size of each file is approximately 500Kbyte in the transmission of images at a time. The problem of the system is to send files using protocol UDP, which may cause error during the transmission, which has an error of about 5 in the 1,000 file in a file that enables DICOM Viewer, a viewer cannot read image files as a whole, because the load will have to check the file before it's error occurred.

### How to solve the problem

The improved cable, may have to change the signal cable from the UDP (User Datagram Protocol) is a fiber optic cable and to modify the program to get files, check the file before the PACS system which allows the speed to slow down while at the medical center want to send data as soon as possible. Medical image system will be connected to the Hospital Information system (HIS) of the medical center, the PACS system will have to write a program to open HIS with Web services and

medical imaging. Overall system, medical center does not want the errors to the solution is: -.

1. Improve the speed of getting information faster.
2. Check the file before importing a DICOM Web Viewer system.
3. Debug the program shuts down when opening images must not come up error dialog box.

### 3. Results and Conclusion

This research has been developed which medical data from MSMC to date ECG and other medical application over wireless board band or Wi-MAX network have been implemented. Such as the ECG signal is estimated to chart a more accurate time series analysis, which we will use the example of the value that occurs over time with equally spaced. The movement of the time series is likely a variation of time series. The trend line is used in the study estimates. Also known as regression analysis. There are different ways such as, logarithmic, polynomial, including the Financial Component, R-squared, the moving average. This research is intended to predict a smooth techniques of ECG signal display. In the first period which will the number of four-wave, such as P-wave, QRS-wave, T-wave and U-wave, so they are to smooth the signal if the value is adjusted to a value that is wrong as shown in Figure 22.




Figure 22. Regression graphs with recession by various means.

Figure 22. A comparison of the signal from the flat with a special exponential moving average and logarithmic functions. In this paper, we use a moving average function to do not put all the data correlation., but It will be compared at a series of n (n is the number of periods to compare data sets).

The method to calculate the moving average the following.

$$S_t = \frac{x_t + x_{t-1} + \dots + x_{t-n+1}}{n}$$

Where is

$S_t$  = forecast at time t.

$X_t$  = the observation time t.

n = number of data used averages (period).




Figure 23. Shows a comparison of the equation moving average signals in different time periods.

Figure 23 can be seen that when increasing the n value of the experimental period by a moving average to smooth the edges even more. But the time period will result in too small waves disappear. In this study, we chose a time period equal to 3 (n = 3). They picked up too much time to make the signal level lower than normal. The signal does not match the actual measurement signal. And when less than 3 periods of time will make the adjustment on uneven. The experimental values obtained from the ECG signal measurement device developed to measure and collect data as a numeric value in every 10 mS by using an application developed.

### 4. Conclusion

With the first year of research funding is still difficult to design systems which have a WiMAX base station delays the import of radio communications, coordination between the organizations, etc. However, researchers have tried to find a solution. But trying to find a solution to the project to continue research. The researcher have tried to research the budget process to ensure maximum benefits. As the first recipient of research grants from the NTC to operate such a project would have difficulty understanding the structure of the high-speed wireless broadband network documenting the progress report in accordance with MOU, etc. This project benefits widely.

## 5.Acknowledgment

The authors also would like to Professor Dr.Prasit Prapinmongkolkarn chairman of Office of National Telecommunications Broadcasting and Telecommunication Commission (NBTC) for his invaluable assistance in funding research. This work was supported by Department of Computer Engineering, Chiang Mai University, Chiang Mai, and Faculty of Medicine, Srinakharinwirot University, Thailand.

## 6.References

- [1] Office of National Telecommunications Broadcasting and Telecommunication Commission (NBTC)'s Medical Application report 2010.
- [2] Planet Communications Asia Co., Ltd. "WiMAX Network Design 2010," Planet communication Confidential Proprifetary. Pp. 46-104.
- [3] K.Noimanee and S.Noimanee, "Medical Consult-Based System for Diagnosis on WiMAX Tchnology," International Journal of Applied Biomedical Engineering, vol.3 No.1 2010. Pp.51-55
- [4] K.Noimanee, S.Noimanee, La-Oopugsin, S.Wattanasirigoonchai. "Development of e-Health Application for Medical Center in National Broadband Project," BMEiCon 4<sup>th</sup> Biomedical International Conference 2011.
- [5] Mustafa Almahdi Algaet, Zul Azri Bin Muhamad Noh, "A Review on Provisioning Quality of Service of Wireless Telemedicine for E-Health Services" Middle-East Journal of Scientific Research 19 (4): 570-592, 2014.